



# **SCOPE OF ACCREDITATION**

Laboratory Name :

POWERGRID ADVANCED RESEARCH AND TECHNOLOGY CENTRE (PARTEC) CALIBRATION LAB, POWERGRID MANESAR COMPLEX, GWALIOR VILLAGE, PACHGAON PO, TAORU ROAD, MANESAR, GURGAON, HARYANA, INDIA

Accreditation Standard Certificate Number Validity

CC-3439 28/07/2022 to 27/07/2024

ISO/IEC 17025:2017

Page No1 of 8Last Amended on-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
		1.0	Permanent Facility	-	_
1	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Current @ (40Hz to 5kHz)	Using 8.5 Digital Multimeter By Direct Method	1 mA to 100 mA	0.06 % to 0.074 %
2	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Current @ (40Hz to 5kHz)	Using 8.5 Digital Multimeter By Direct Method	100 mA to 2 A	0.074 % to 0.15 %
3	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Current @ (40Hz to 5kHz)	Using 8.5 Digital Multimeter By Direct Method	2 A to 20 A	0.15 % to 0.7 %
4	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Voltage @ (10Hz to 45Hz)	Using 8.5 Digital Multimeter By Direct Method	1 mV to 200 mV	2.12 % to 0.2 %





## **SCOPE OF ACCREDITATION**

Laboratory Name :

POWERGRID ADVANCED RESEARCH AND TECHNOLOGY CENTRE (PARTEC) CALIBRATION LAB, POWERGRID MANESAR COMPLEX, GWALIOR VILLAGE, PACHGAON PO, TAORU ROAD, MANESAR, GURGAON, HARYANA, INDIA

Accreditation Standard Certificate Number Validity ISO/IEC 17025:2017 CC-3439 28/07/2022 to 27/07/2024

Last Amended on

2 of 8

\_

Page No

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Voltage @ (10Hz to 45Hz)	Using 8.5 Digital Multimeter By Direct Method	200 mV to 20 V	0.2 % to 0.26 %
6	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Voltage @ (10kHz to 20kHz)	Using 8.5 Digital Multimeter By Direct Method	1 mV to 200 mV	1 % to 0.4 %
7	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Voltage @ (10kHz to 20kHz)	Using 8.5 Digital Multimeter By Direct Method	200 mV to 200 V	0.4 % to 0.034 %
8	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Voltage @ (45Hz to 10kHz)	Using 8.5 Digital Multimeter By Direct Method	1 mV to 200 mV	0.7 % to 0.02 %
9	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Voltage @ (45Hz to 10kHz)	Using 8.5 Digital Multimeter By Direct Method	1 V to 1000 V	0.029 % to 0.18 %





# **SCOPE OF ACCREDITATION**

Laboratory Name :

POWERGRID ADVANCED RESEARCH AND TECHNOLOGY CENTRE (PARTEC) CALIBRATION LAB, POWERGRID MANESAR COMPLEX, GWALIOR VILLAGE, PACHGAON PO, TAORU ROAD, MANESAR, GURGAON, HARYANA, INDIA

Accreditation Standard Certificate Number Validity ISO/IEC 17025:2017 CC-3439

28/07/2022 to 27/07/2024

Page No3 of 8Last Amended on-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
10	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Measure)	AC Voltage @ (45Hz to 10kHz)	Using 8.5 Digital Multimeter By Direct Method	200 mV to 1 V	0.02 % to 0.029 %
11	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current @ (45Hz to 1kHz)	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 A to 10 A	0.064 % to 0.11 %
12	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current @ (45Hz to 1kHz)	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 mA to 100 mA	0.11 % to 0.048 %
13	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Current @ (45Hz to 1kHz)	Using 6.5 Digit Multifunction Calibrator By Direct Method	100 mA to 1 A	0.048%
14	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ (45 Hz to 10kHz)	Using 6.5 Digit Multifunction Calibrator By Direct Method	2 mV to 1 V	0.3 % to 0.03 %





## **SCOPE OF ACCREDITATION**

Laboratory Name :

POWERGRID ADVANCED RESEARCH AND TECHNOLOGY CENTRE (PARTEC) CALIBRATION LAB, POWERGRID MANESAR COMPLEX, GWALIOR VILLAGE, PACHGAON PO, TAORU ROAD, MANESAR, GURGAON, HARYANA, INDIA

Accreditation Standard Certificate Number Validity ISO/IEC 17025:2017 CC-3439

28/07/2022 to 27/07/2024

Last Amended on

Page No

4 of 8

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
15	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ (45Hz to 10kHz)	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 V to 100 V	0.03 % to 0.02 %
16	ELECTRO- TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage @ (45Hz to 10kHz)	Using 6.5 Digit Multifunction Calibrator By Direct Method	100 V to 1000 V	0.02 % to 0.03 %
17	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Current	Using 8.5 Digital Multimeter By Direct Method	1 μA to 100 μA	0.7 % to 0.7 %
18	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Current	Using 8.5 Digital Multimeter By Direct Method	1 A to 20 A	0.02 % to 0.046 %
19	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Current	Using 8.5 Digital Multimeter By Direct Method	100 µA to 1 A	0.7 % to 0.02 %
20	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Resistance	Using 8.5 Digital Multimeter By Direct Method	1 kohm to 100 kohm	0.015%





## **SCOPE OF ACCREDITATION**

Laboratory Name : Accreditation Standard Certificate Number

Validity

POWERGRID ADVANCED RESEARCH AND TECHNOLOGY CENTRE (PARTEC) CALIBRATION LAB, POWERGRID MANESAR COMPLEX, GWALIOR VILLAGE, PACHGAON PO, TAORU ROAD, MANESAR, GURGAON, HARYANA, INDIA

CC-3439 28/07/2022 to 27/07/2024

ISO/IEC 17025:2017

Page No5 of 8Last Amended on-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
21	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Resistance	Using 8.5 Digital Multimeter By Direct Method	1 ohm to 1 kohm	0.015 % to 0.015 %
22	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Resistance	Using 8.5 Digital Multimeter By Direct Method	100 kohm to 100 Mohm	0.015 % to 0.03 %
23	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Resistance	Using 8.5 Digital Multimeter By Direct Method	100 Mohm to 1 Gohm	0.03 % to 1.3 %
24	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Voltage	Using 8.5 Digital Multimeter By Direct Method	1 mV to 100 mV	0.02 % to 0.002 %
25	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Voltage	Using 8.5 Digital Multimeter By Direct Method	10 V to 1000 V	0.002 % to 0.001 %
26	ELECTRO- TECHNICAL- DIRECT CURRENT (Measure)	DC Voltage	Using 8.5 Digital Multimeter By Direct Method	100 mV to 10 V	0.002%





## **SCOPE OF ACCREDITATION**

Laboratory Name :

POWERGRID ADVANCED RESEARCH AND TECHNOLOGY CENTRE (PARTEC) CALIBRATION LAB, POWERGRID MANESAR COMPLEX, GWALIOR VILLAGE, PACHGAON PO, TAORU ROAD, MANESAR, GURGAON, HARYANA, INDIA

Accreditation Standard Certificate Number Validity ISO/IEC 17025:2017 CC-3439

28/07/2022 to 27/07/2024

Last Amended on

Page No

6 of 8

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
27	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Capacitance	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 nF to 100 nF	0.212 % to 1.14 %
28	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Capacitance	Using 6.5 Digit Multifunction Calibrator By Direct Method	100 nF to 1 µF	0.212 % to 0.275 %
29	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Current	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 μA to 300 μA	1.6 % to 0.2 %
30	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Current	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 A to 10 A	0.02 % to 0.05 %
31	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Current	Using 6.5 Digit Multifunction Calibrator By Direct Method	30 mA to 1 A	0.09 % to 0.02 %
32	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Current	Using 6.5 Digit Multifunction Calibrator By Direct Method	300 μA to 30 mA	0.2 % to 0.09 %





## **SCOPE OF ACCREDITATION**

Laboratory Name :

POWERGRID ADVANCED RESEARCH AND TECHNOLOGY CENTRE (PARTEC) CALIBRATION LAB, POWERGRID MANESAR COMPLEX, GWALIOR VILLAGE, PACHGAON PO, TAORU ROAD, MANESAR, GURGAON, HARYANA, INDIA

**Accreditation Standard Certificate Number** Validity

ISO/IEC 17025:2017 CC-3439 28/07/2022 to 27/07/2024

**Measurand or Reference** 

Material/Type of instrument

Page No Last Amended on 7 of 8

Measurement range and \* Calibration and additional parameters Measurement

S.No	Discipline / Group	Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
33	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Resistance	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 kohm to 300 kohm	0.004 % to 0.007 %
34	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Resistance	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 Mohm to 300 Mohm	0.0032 % to 0.25 %
35	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Resistance	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 Ohm to 1 kohm	0.8 % to 0.004 %
36	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Resistance	Using 6.5 Digit Multifunction Calibrator By Direct Method	300 kohm to 1 Mohm	0.0049 % to 0.0032 %
37	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Resistance	Using 6.5 Digit Multifunction Calibrator By Direct Method	300 Mohm to 1000 Mohm	0.25 % to 1.177 %
38	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	Using 6.5 Digit Multifunction Calibrator By Direct Method	1 mV to 100 mV	0.08 % to 0.003 %





## **SCOPE OF ACCREDITATION**

Laboratory Name :

POWERGRID ADVANCED RESEARCH AND TECHNOLOGY CENTRE (PARTEC) CALIBRATION LAB, POWERGRID MANESAR COMPLEX, GWALIOR VILLAGE, PACHGAON PO, TAORU ROAD, MANESAR, GURGAON, HARYANA, INDIA

Accreditation Standard Certificate Number Validity

CC-3439 28/07/2022 to 27/07/2024

ISO/IEC 17025:2017

Page No Last Amended on 8 of 8

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
39	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	Using 6.5 Digit Multifunction Calibrator By Direct Method	10 V to 1000 V	0.002 % to 0.0016 %
40	ELECTRO- TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	Using 6.5 Digit Multifunction Calibrator By Direct Method	100 mV to 10 V	0.003 % to 0.002 %
41	ELECTRO- TECHNICAL- TEMPERATURE SIMULATION (Source)	Temperature Simulation RTD Pt 385	Using 6.5 Digit Multifunction Calibrator By Direct Method	-200 °C to 600 °C	0.091°C
42	ELECTRO- TECHNICAL- TIME & FREQUENCY (Source)	Frequency	Using 6.5 Digit Multifunction Calibrator By Direct Method	10 Hz to 100 KHz	0.0008 % to 0.0002 %

\* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.